

MD 100 Photometer

Precise Water Analysis in High-Quality Design

Small | Mobile | Rapid

The MD 100 uses high quality interference filters with long-life LEDs as a light source without any moving parts in a transparent sample chamber.

The units supply accurate, reproducible results very quickly. Other major advantages include ease of operation, ergonomic design, compact dimensions and safe handling.

The calibration and software-based adjustment options mean that the MD 100 is also suitable for use as a testing instrument.

The tests are conducted using either Lovibond® tablet reagents with long-term stability and a guaranteed minimum 5 or 10 year shelf life, VARIO powder reagents or using liquid reagents.

▶ Please see pages 78 onwards for reagents (order codes)

Highlights

- Scroll memory
- Automatic switch-off
- Real-Time-Clock and date
- Calibration mode
- Backlit display
- Storage function
- One Time Zero (OTZ)
- Waterproof*)

*) as defined in IP 68, 1 hour at 0.1 meter



Single-Parameter

| Test | Code |
|---|----------|
| Aluminium , tablet reagents 0.01 - 0.3 mg/l Al | 27 62 00 |
| Aluminium , powder reagents 0.01 - 0.25 mg/l Al | 27 62 05 |
| Ammonia , tablet reagents 0.02 - 1.0 mg/l N | 27 60 60 |
| Ammonium , powder reagents 0.01 - 0.8 mg/l N | 27 60 65 |
| Chloride , tablet reagents 0.5 - 25 mg/l Cl ⁻ 5 - 250 mg/l Cl ⁻ (by dilution) | 27 61 80 |
| Chlorine , tablet reagents (OTZ) 0.01 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * | 27 60 00 |
| Chlorine , liquid reagent (OTZ) 0.02 - 4 mg/l Cl ₂ | 27 60 05 |
| Chlorine DUO , for 2 types of reagents 1) Tablet reagents 0.01 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * | 27 60 20 |
| 2) Powder reagents 0.02 - 2.0 mg/l Cl ₂ (ø 24 mm glass vial) 0.1 - 8.0 mg/l Cl ₂ (ø 10 mm multi vial-2) | 27 60 25 |
| Chlorine , powder reagents 0.02 - 2.0 mg/l Cl ₂ (ø 24 mm glass vial) 0.1 - 8.0 mg/l Cl ₂ (ø 10 mm multi vial-2) | 27 60 10 |
| Chlorine HR (Potassium iodide) tablet reagents 5 - 200 mg/l Cl ₂ (ø 16 mm round vial & adapter) | 27 61 70 |
| Chlorine dioxide , tablet reagents 0.02 - 11 mg/l ClO ₂ | 27 60 30 |
| Chlorine dioxide , powder reagents 0.04 - 3.8 mg/l ClO ₂ | 27 60 35 |
| COD , tube tests, without reagents 0 - 150 mg/l O ₂ (ø 16 mm) 0 - 1500 mg/l O ₂ (ø 16 mm) 0 - 15000 mg/l O ₂ (ø 16 mm) | 27 61 20 |
| Copper , tablet reagents 0.05 - 5.0 mg/l Cu | 27 60 80 |
| Copper , powder reagents 0.05 - 5.0 mg/l Cu | 27 60 85 |
| Hardness, total , tablet reagents 2 - 50 mg/l CaCO ₃ 20 - 500 mg/l CaCO ₃ (by dilution) | 27 61 90 |
| Hazen , no reagents required 0 - 500 mg/l Pt-Co | 27 61 60 |
| Iron , tablet reagents 0.02 - 1.0 mg/l Fe | 27 60 50 |
| Iron TPTZ , powder reagents 0.02 - 1.8 mg/l Fe | 27 60 55 |
| Iron , powder reagents 0.02 - 3.0 mg/l Fe | 27 60 56 |
| Fluoride , without reagents 0.05 - 2.0 mg/l F ⁻ | 27 60 90 |
| Manganese LR , tablet reagents 0.2 - 4.0 mg/l Mn | 27 61 00 |
| Manganese LR , powder reagents 0.01 - 0.7 mg/l Mn | 27 61 05 |
| Manganese HR , powder reagents 0.1 - 18 mg/l Mn | 27 61 06 |
| Molybdenum LR Powder reagents / reagent solution 0.03 - 3.0 mg/l Mo (mixing cylinder required, not included) | 27 61 40 |
| Molybdenum HR , powder reagents 0.3 - 40 mg/l Mo | 27 61 41 |
| Molybdenum , tablet reagents 0.6 - 30 mg/l Mo | 27 61 42 |

Single-Parameter

| Test | Code |
|---|----------|
| Phosphate , tablet reagents 0.05 - 4.0 mg/l PO ₄ | 27 60 40 |
| Phosphate , powder reagents 0.06 - 2.5 mg/l PO ₄ | 27 60 45 |
| Silica , tablet reagents 0.05 - 4.0 mg/l SiO ₂ | 27 61 10 |
| Silica LR , powder reagents 0.1 - 1.6 mg/l SiO ₂ | 27 61 15 |
| Silica HR , powder reagents 1 - 90 mg/l SiO ₂ | 27 61 16 |
| Suspended solids no reagents required 0 - 750 mg/l TSS | 27 61 50 |
| Urea , tablet reagents 0.1 - 2.5 mg/l Urea 0.2 - 5 mg/l Urea (by dilution) | 27 62 10 |

2in1

| | |
|--|----------|
| Chlorine, pH , tablet reagents (OTZ) 0.01 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * 6.5 - 8.4 pH | 27 80 20 |
| Chlorine, pH , liquid reagent (OTZ) 0.02 - 4 mg/l Cl ₂ / 6.5 - 8.4 pH | 27 80 25 |
| Chlorine, pH , powder reagents for chlorine 0.02 - 2.0 mg/l Cl ₂ (ø 24 mm glass vial) 0.1 - 8.0 mg/l Cl ₂ (ø 10 mm multi vial-2) 6.5 - 8.4 pH | 27 80 30 |

3in1

| | |
|---|----------|
| Chlorine, pH, Cyanuric acid tablet reagents (OTZ) 0.01 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * 6.5 - 8.4 pH ; 0 - 160 mg/l cyanuric acid | 27 80 10 |
| Chlorine, pH, Cyanuric acid liquid reagent for chlorine and pH (OTZ) 0.02 - 4 mg/l Cl ₂ / 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid | 27 80 15 |
| Chlorine, pH, Alkalinity-M tablet reagents (OTZ) 0.01 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * 6.5 - 8.4 pH ; 5 - 200 mg/l CaCO ₃ (TA) | 27 80 60 |
| Chlorine, pH, Alkalinity-M (total) liquid reagent for chlorine and pH (OTZ) 0.02 - 4 mg/l Cl ₂ / 6.5 - 8.4 pH 5 - 200 mg/l CaCO ₃ (TA) | 27 80 65 |
| Chlorine LR, Chlorine HR, Chlorine dioxide [#] , tablet reagents 0.01 - 6.0 mg/l Cl ₂ 5 - 200 mg/l Cl ₂ (ø 16 mm round vial) 0.02 - 11 mg/l ClO ₂ | 27 80 00 |

4in1

| Test | Code |
|---|----------|
| Chlorine, pH, Cyanuric acid, Alkalinity-M , tablet reagents (OTZ) 0.02 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * 6.5 - 8.4 pH ; 0 - 160 mg/l cyanuric acid 5 - 200 mg/l CaCO ₃ (TA) | 27 80 70 |
| Chlorine, pH, Cyanuric acid, Alkalinity-M (total) liquid reagent for chlorine and pH (OTZ) 0.02 - 4 mg/l Cl ₂ / 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid / 5 - 200 mg/l CaCO ₃ (TA) | 27 80 75 |

5in1

| | |
|---|----------|
| Chlorine, pH, Cyanuric acid, Alkalinity-M, Calcium hardness tablet reagents (OTZ) 0.02 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * 6.5 - 8.4 pH ; 0 - 160 mg/l cyanuric acid 5 - 200 mg/l CaCO ₃ (TA) ; 0 - 500 mg/l CaCO ₃ (CaH) | 27 80 80 |
|---|----------|

6in1

| | |
|--|----------|
| Chlorine, Bromine, pH, Cyanuric acid, Alkalinity-M, Calcium hardness , tablet reagents (OTZ) 0.02 - 6.0 mg/l Cl ₂ / 0.1 - 10 mg/l Cl ₂ * 0.05 - 13 mg/l Br ; 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid ; 5 - 200 mg/l CaCO ₃ (TA) 0 - 500 mg/l CaCO ₃ (CaH) | 27 80 90 |
|--|----------|

MD 100 Boiler Water

| | |
|---|----------|
| Aluminium, Chloride, Copper, DEHA, Hydrazine, Iron, Oxygen (dissolved), Phosphate, Polyacrylate, Silica (delivery without reagents) | 27 62 30 |
|---|----------|

MD 100 Cooling Water

| | |
|---|----------|
| Aluminium, Bromine, Chlorine, Chlorine HR, Chlorine dioxide, Copper, Iron, Iron in Mo, Molybdate LR, Molybdate HR, Ozone, Polyacrylate, Sulphate, Triazoles, Zinc (delivery without reagents) | 27 62 40 |
|---|----------|

* Delivery without reagents
for measuring range 0.1 - 10 mg/l Cl₂

Where chlorine and chlorine dioxide are present together,
they may be determined quantitatively as a single figure.

MD 100 Photometer



Technical Data

| | |
|--|---|
| Optics | LEDs, interference filters (IF) and photo sensor in transparent sample chamber. Depending on the version, up to 3 different interference filters are used. Wavelength specifications of interference filters: 430 nm $\Delta\lambda = 5$ nm 530 nm $\Delta\lambda = 5$ nm 560 nm $\Delta\lambda = 5$ nm 580 nm $\Delta\lambda = 5$ nm 610 nm $\Delta\lambda = 6$ nm 660 nm $\Delta\lambda = 5$ nm |
| Wavelength Accuracy | ± 1 nm |
| Photometric Accuracy⁴⁾ | 3 % FS (T = 20 °C – 25 °C) |
| Photometric Resolution | 0.01 A |
| Power Supply | 4 micro batteries (AAA), capacity approx. 17 hours or 5000 tests |
| Auto - OFF | automatic switch-off |
| Display | backlit LCD (on keypress) |
| Storage | internal ring memory for 16 data sets |
| Interfaces | infrared interface for test data transfer |
| Additional feature | real time clock and date |
| Calibration | factory calibration and user calibration. Reset to factory calibration possible |
| Dimensions | 155 x 75 x 35 mm (L x W x H) |
| Weight | basic unit approx. 260 g |
| Environmental conditions | temperature: 5–40 °C rel. humidity: 30–90 % (non condensing) |

CE-Conformity

⁴⁾ tested with standard solutions

Scroll Memory (SM)

To avoid unnecessary scrolling for the required test method, the instrument memorizes the last method used before switching off the instrument. When the instrument is switched on again, the scroll list comes up with the last used test method first.

Zero Setting (OTZ)

For certain versions of the instrument it is not necessary to zero the instrument each time. The zero setting is held in memory until the device is turned off. (One Time Zero - OTZ). The zero setting can be confirmed whenever it is required.

Manufacturers Test Certificate M

Besides the "Certificate of Compliance" which is supplied with the MD 100, manufacturers test certificates M are available at cost on request. Manufacturers test certificates M are individually supplied per instrument and per method.

The manufacturers test certificate M has to be ordered together with the new instrument and cannot be delivered at a later stage.

N.I.S.T Traceability

The instrument has a factory calibration, which is related to international standards which are not N.I.S.T traceable. The instrument may be calibrated by the user in a "user calibration mode" with N.I.S.T traceable standards.

(N.I.S.T. = National Institute of Standards and Technology)

Delivery Content

- Instrument in carrying case
- 4 micro batteries (AAA)
- 3 Round vials (glass) with lid
- 1 stirring rod & 1 brush
- Tablet reagents and/or liquid reagents or VARIO Powder reagent
- Warranty information
- Certificate (COC)
- Instruction Manual



Accessories

| Item | Code |
|---|-------------|
| Set of 12 round vials with lid Height 48 mm, Ø 24 mm | 19 76 20 |
| Set of 5 round vials with lid Height 48 mm, Ø 24 mm | 19 76 29 |
| Set of 10 round vials with lid Height 90 mm, Ø 16 mm | 19 76 65 |
| Adapter for round vials Ø 16 mm | 19 80 21 90 |
| Set of 12 plastic vials (PC), with lid "Multi"-Type 2 , Ø 10 mm | 19 76 00 |
| Vial stand for 6 round vials Ø 24 mm, acrylic glass | 41 89 51 |
| Vial stand for 10 vials (Ø 16 mm or □ 13,5 mm), acrylic glass | 41 89 57 |
| Mixing cylinder, 25 ml, with stopper required accessory for molybdenum LR test with MD 100 (276140) | 19 80 26 50 |
| Membrane filter set for use when preparing samples, 25 membrane filters, 0,45 µm, 2 syringes 20 ml | 36 61 50 |
| Cleaning cloth for vials | 19 76 35 |
| Set of 12 sealing rings for round vial Ø 24 mm | 19 76 26 |
| 4 micro batteries (AAA) | 19 50 026 |
| Measuring beaker, volume 100 ml | 38 48 01 |
| Plastic funnel with handle | 47 10 07 |
| Plastic stirring rod, 13 cm length | 36 41 00 |
| Plastic stirring rod, 13 cm length, (10 pc.) | 36 41 20 |
| Plastic stirring rod, 10 cm length | 36 41 09 |
| Plastic stirring rod, 10 cm length, (10 pc.) | 36 41 30 |
| Infrared data transfer module IRiM | 21 40 50 |



 Please see pages 78 onwards for reagents (order codes)



Data transfer

The optional available IRiM (infrared interface module) uses modern infrared technology to transmit measurement data from the MD 100 photometer to one of 3 optional interfaces. These interfaces can be used to connect to a PC, a USB printer¹⁾ or alternatively a serial printer²⁾.

The unit is supplied complete with data logging software providing easy and rapid transfer of data to the PC. As an option, the data can be saved as an Excel sheet or a .txt file.

Measurement data can quickly be printed out, using a specified¹⁾ USB or alternatively a printer with a serial plug-in connected to the IRiM.

Applicable for the following operating systems: Windows® XP, Windows® Vista and Windows® 7/10.

¹⁾ USB printer: HP Deskjet 6940 ; ²⁾ each ASCII printer
Windows® is a registered Trademark of Microsoft Corporation

Verification Standard Kit

The verification standard kit for the MD 100 is designed to assure the user of the accuracy and the reliability of the results related to the integrated wave lengths.

The kit contains one zero standard, 6 different vials for checking 6 different wave lengths and allows checking the complete range of MD 100 photometers.

The shelf life of the verification standard kit is two years from the date of production, provided that storage and use are in accordance with the instructions provided.

Verification Standard Kit 21 56 70

Reference Standard Kit for MD 100

The reference standards are designed to check the accuracy and the reliability of the results.

It is not possible to calibrate the photometer with the reference standards.

The shelf life of reference standards is two years from the date of production, provided that storage and use are in accordance with the instructions provided.

Kit Chlorine for instruments with tablet / liquid reagent 0.2* and 1.0* mg/l 27 56 50

Kit Chlorine for instruments with tablet / liquid reagent 0.5* and 2.0* mg/l 27 56 55

Kit Chlorine for instruments with tablet / liquid reagent 1.0* and 4.0* mg/l 27 56 56

Kit Chlorine for instruments with powder reagent (VARIO) 0.2* and 1.0* mg/l 27 56 60

Kit pH for instruments with tablet / liquid reagent 7,45* pH 27 56 70

* Approximate figure, actual figure specified in Certificate of Analysis

